

## 0-1" / 0-25mm Electronic Depth / Pit Gauge

Cat # SPI13-749-7



- Large LCD - 0.0005" (0.01mm) Resolution
- $\pm 0.001$ " (0.025mm) Accuracy
- 0.06" Dia Probes
- 2 3/8" Base
- Fitted Case & Battery Included
- Certificate of Calibration Traceable to NIST

*Measurement Scale: Inch / Metric*



G.A.L. Gage Company

Phone: (269) 465-5750 • Fax: (269) 465-6385 • [www.galgage.com](http://www.galgage.com)  
2953 Hinchman Road, P.O. Box 218, Stevensville, Michigan 49127

